

Abstract

There is described a method for controlling the power provided by a power supply to magnetic-inductive flow measurement devices. The instantaneous signal-to-noise ratio is determined and the power provided to the flow measurement device is controlled to be inversely proportional to that ratio so that as that ratio increases the power decreases and as that ratio decreases the power increases.